

High Sensitivity C-Reactive Protein (hs-CRP) Turbidimetric Immunoassay Kit

Catalogue number: 51121

For the quantitative determination of C-reactive Protein
in human serum and plasma

This package insert must be read in its entirety before using this product
Use only the current version of product data sheet enclosed with the kit

Website: www.immunodiagnostics.com.hk

E-mail: info@immunodiagnostics.com.hk

Tel: (+852) 3502 2780

Fax: (+852) 3502 2781

**FOR RESEARCH USE ONLY
NOT FOR USE IN DIAGNOSTIC PROCEDURES**

Version: 3.1



TABLE OF CONTENT

Content	Page
INTRODUCTION	1
PRINCIPLE OF THE ASSAY	1
REAGENTS SUPPLIED	2
OTHER MATERIALS REQUIRED, BUT NOT PROVIDED	2
STORAGE	2
SAMPLE HANDLING	2
ASSAY PROCEDURE	3
TYPICAL STANDARD CURVE	3
CALCULATION	4
ASSAY CHARACTERISTICS	4

PACKING SPECIFICATION

Cat. No.	Size	Approximately tests
51121-05	R1: 15ml, R2: 5ml	100
51121-10	R1: 30ml, R2: 10ml	200
51121-20	R1: 60ml, R2: 20ml	400
51121-50	R1: 150ml, R2: 50ml	1000
51121-100	R1: 300ml, R2: 100ml	2000

INTRODUCTION

C-Reactive Protein (CRP), also known as PTX1, and pentraxin-related, is a sensitive biomarker for the early diagnostics and observation of curative effect of the liver diseases and malnutrition. Because it is mainly synthesized by hepatic cells, and its serum levels decreased significantly under the condition of liver injury or malnutrition.

Relatively high levels of hs-CRP in otherwise healthy individuals have been found to be predictive of an increased risk of a future heart attack, stroke, sudden cardiac death, and/or peripheral arterial disease, even when cholesterol levels are within an acceptable range.

People with higher hs-CRP values have the highest risk of cardiovascular disease and those with lower values have less risk. Specifically, individuals who have hs-CRP results at the high end of the normal range have 1.5 to 4 times the risk of having a heart attack as those with hs-CRP values at the low end of the normal range.

The American Heart Association and U.S. Centers for Disease Control and Prevention have defined risk groups as follows: Low risk: less than 1.0 mg/L; Average risk: 1.0 to 3.0 mg/L; High risk: above 3.0 mg/L

PRINCIPLE OF THE ASSAY

This assay is a turbidimetric immunoassay for the quantitative measurement of CRP in human serum and plasma. A standard or sample is added into a cuvette and mixed with the reaction buffer R1. After a short incubation, the test reagent R2, which is a suspension of microparticles coated with CRP antibodies, is added into the cuvette and mixed. The presence of CRP in the standard or sample causes the immune-particles to aggregate. The extent to which the microparticles

aggregate is quantified by the amount of light scattering measured as absorbance by a chemistry analyzer. The concentration of CRP in unknown samples can be interpolated from a reference curve using the standards provided.

REAGENTS SUPPLIED

R1 – Reaction buffer, 30 ml, a ready-to-use buffer solution containing salt, polyether compound and preservative

R2 – Test reagent, 10 ml, a ready-to-use suspension of polymer microparticles coated with rabbit anti-ADPN polyclonal antibodies in storage buffer

OTHER MATERIALS REQUIRED, BUT NOT PROVIDED

1. Clinical chemistry analyzer
2. CRP calibrators (provided separately, Cat. #51121-S1)
3. CRP controls (optional, provided separately, Cat. #51121-C1)
4. Deionized water
5. Analyzer-specific reagent containers for R1 and R2

STORAGE

The kit should be stored at 2-8°C upon receipt. Once opened, the reagents may be stored at 2-8°C for up to 4 weeks.

SAMPLE HANDLING

This kit can be used to determine CRP in human serum and plasma samples. Blood specimens should be collected aseptically into appropriate tubes. Plasma should be prepared by standard techniques for laboratory testing. The prepared specimens should be stored in closed vessels. If the assay cannot be performed within 24 hours or specimens are to be shipped, the specimens should be frozen at -20°C or below. For long-term storage of specimens, -70°C or below is recommended. To avoid freeze-thaw cycles, specimens should be aliquoted. Do not use hemolyzed, hyperlipemic, heat-treated or contaminated specimens. No dilution of the sample is required in this assay.

ASSAY PROCEDURE


Assay procedures may vary depending on the automated chemistry analyzer to be used. A general example of assay procedures is stated as follow:

1. Dispense 300µl of R1 into a clean cuvette
2. Add 3µl of sample and incubate at 37°C for 5 minutes
3. Further add 100µl of R2
4. Read change of absorbance at Main Wavelength 570 nm and Sub Wavelength 800 nm for 8 minutes after the addition of R2
5. Calculate the concentration of CRP in unknown sample by interpolation from a reference curve using the standards provided

TYPICAL STANDARD CURVE

The following standard curve is provided for demonstration only. A standard curve should be generated for each assay.

CRP (mg/L)	Absorbance (570 nm)
0	0
0.5	0.098
2	0.361
8	1.468
20	3.144
50	4.845



ADPN (mg/L)
3

CALCULATION

1. Subtract the absorbance of the blank from that of standards and samples.
2. Generate a standard curve by plotting the absorbance obtained (y-axis) against CRP concentrations (x-axis). The best fit line can be generated with any curve-fitting software by regression analysis. 4-parameter curve fitting can be used for calculation.
3. Determine CRP concentration of samples from standard curve.

ASSAY CHARACTERISTICS

A. Sensitivity

The sensitivity is defined as the lower limit of detection and is estimated as the mean of the blank sample plus three times the SD obtained from the blank sample. The sensitivity of CRP assay is 0.2 mg/L.

B. Precision

The precision of the CRP assay is < 10% CV. Four samples consisting of two CRP controls and two serum based panels were assayed 20 times separately.

Sample	Mean CRP (mg/L)	SD (mg/L)	CV
Low Control	0.85	0.3	4.0%
High Control	3.11	0.2	3.4%
Panel 1	3.62	0.09	2.4%
Panel 2	1.75	0.05	2.6 %

C. Linearity

The CRP assay is linear between 0.2 mg/L to 50 mg/L.

D. Interference

No interference was detected with hemoglobin up to 5 g/L, conjugated bilirubin up to 300 mg/L, free bilirubin up to 300 mg/L, and up to 5g/L lipid emulsion.



5F, Biotech Centre 2 (11W), No. 11 Science Park
West Avenue, Hong Kong Science Park, Shatin. NT, Hong Kong
Website: www.immunodiagnostics.com.hk
E-mail: sales@immunodiagnostics.com.hk
Tel: (+852) 3502 2780
Fax: (+852) 3502 2781